



Servicebio

产品检验报告

Product Test Report

| | | | |
|----------------------|-------------------------------------|---------------------------|---------------------------|
| 产品名称 Product name | 离心管 (PCR) Centrifuge tube(PCR) | 产品材质 Material | PP (聚丙烯) Polypropylene |
| 产品型号 Product mode | 排盖 8 连 8-strip tube with row cap | 产品规格 Specification | 0.2 mL (200 μ l) |
| 产品批号 Lot number | IMC20211119 (管盖: IMC20211119) | 产品货号 Product number | PCR-0802-FC |
| 产品状态 Status | 无酶 Dnase/Rnase free | 储存条件 Storage condition | 常温 Room temperature |

检验依据: 《中华人民共和国药典》2020 版、GB/T 2828.1-2012

Test basis: Pharmacopoeia of the People's Republic of China 2020 edition, GB/T 2828.1-2012

抽样原则: GB/T 2828.1-2012; 外观性能的接收质量限 AQL=0.40、其它接收质量限 AQL=0.01

Sampling principle: GB/T 2828.1-2012; The receiving quality limit of appearance performance is AQL=0.40, and the other receiving quality limit is AQL=0.01.

检验项目 Test items

| 检验指标 Test index | | 检验标准 Test standard | 检验结果 Test result | 判定 Decision |
|--------------------|------------------|--|--|-----------------|
| 外观 Appearance | 毛刺 Burrs | 离心管口、盖和管体有长于 0.5mm 的毛刺均属于不合格品; The burr longer than 0.5mm on the nozzle, cover and tube body on the centrifuge tube is unqualified. | 所有规格均合格 All the products are qualified | 合格 Qualified |
| | 缺损 Defect | 产品有任何部分的缺损均属于不合格品。 Any defect in any part of the product is unqualified. | 所有规格均无缺损 All the products are free of defects | 合格 Qualified |
| | 变形 Distortion | 产品表面的各种缺损如开裂、扭曲变形、污点以及肉眼可见 (日光灯下距离 30 厘米) 气泡均属于不合格产品。All kinds of defects on the surface of the product, such as cracking, distortion, stains and visible to the naked eye (30cm away from the fluorescent lamp) bubble are unqualified products. | 所有规格均无变化 All the products are free of distortion | 合格 Qualified |
| | 污染 Contamination | 产品有类似油脂等物质的均属于不合格品; 黑点: 产品有肉眼可见 (日光灯下距离 30 厘米) 的小黑点属于不合格产品。Products with substances similar to oils and fats belong to unqualified products; Black spot: the product has the naked eye visible small black spot (fluorescent lamp distance 30cm) belongs to unqualified product. | 所有规格均无污染 All specifications are free of contamination | 合格 Qualified |
| | 内陷 Invagination | 产品有任何肉眼可见 (日光灯下距离 40 厘米) 的内陷均属于不合格品。Product has the naked eye visible invagination (fluorescent lamp distance 40cm) belongs to unqualified product. | 所有规格均无内陷 All specifications are free of invagination | 合格 Qualified |



Servicebio

| | | | |
|---|--|-----------------|-----------------|
| RNA 酶检验 RNA enzyme test | 荧光定量分析法：检测样品均为阴性 Fluorescence quantitative analysis: all the samples tested were negative. | 阴性 Negative | 合格 Qualified |
| DNA 酶检验 DNA enzyme test | 荧光定量分析法：检测样品均为阴性 Fluorescence quantitative analysis: all the samples tested were negative. | 阴性 Negative | 合格 Qualified |
| 内毒素检验 Endotoxin test | 《中华人民共和国药典》2020 版小于 0.1EU/mL：检测样品均阴性 <i>Pharmacopoeia of the People's Republic of China</i> 2020 version less than 0.1EU/mL: All the tested samples were negative. | 阴性 Negative | 合格 Qualified |
| 化学相容性 Chemical compatibility | 本产品能耐硫酸、硝酸、盐酸、氢氧化钠（70%）、过氧化氢（90%）等化学试剂 This product can resist sulfuric acid, nitric acid, hydrochloric acid, sodium hydroxide (70%), hydrogen peroxide (90%) and other chemical reagents. | 耐受 Tolerance | 合格 Qualified |
| <p>检验结果：</p> <p>根据产品检验参考标准中检验项目进行检验，所有检验项目均符合要求，因此此批产品检验合格，可入库发货。</p> <p>Test result:</p> <p>According to the inspection items in the product inspection reference standard, all inspection items meet the requirements. Therefore, this batch of products can be stored and delivered.</p> <div style="text-align: center;">  <p>质检专用章 Special stamp for quality inspection</p> <p>签发日期：2021 年 11 月 19 日</p> <p>Issuing date: 2021.11.19</p> </div> | | | |

检验：王滴、桂良凤 审核：李子庆 批准：刘小影 表单编号：SWE-QA-SMP-006-03

Inspected by: Wang Di, Gui Liangfeng Checked by: Li Ziqing Approved by: Liu Xiaoying No: SWE-QA-SMP-006-03