



GenElute™ UltraMag Cell-Free DNA Kit

Catalog No. CFMAG

FOR RESEARCH USE ONLY
Not for use in diagnostic procedures.

USA & Canada Phone: +1(800) 645-5476

In Europe, please contact Customer Service:

France: 0825.045.645;

Spain: 901.516.645 Option 1

Germany: 01805.045.645

Italy: 848.845.645

United Kingdom: 0870.900.46.45

For other locations across the world please visit www.millipore.com/offices

Introduction

GenElute™ UltraMag Cell-Free DNA Kit provides rapid and efficient purification of circulating free DNA, also known as cfDNA. The kit uses silica-coated magnetic beads to purify cell free DNA from serum and plasma samples of less than 1ml up to 10mL. The kit was developed to provide efficient recovery cell-free DNA fragments in the range of 100 bp – 500 bp. The recovered DNA is suitable for a wide range of down-stream applications including next-generation sequencing, qPCR and bisulfite sequencing.

For Research Use Only; Not for use in diagnostic procedures

Product Overview

GenElute™ UltraMag Cell-Free DNA Kit allows for fast and efficient cell-free DNA (cfDNA) isolation from plasma/serum samples. The magnetic bead-based extraction protocol is ideally suited for use with robotic liquid handlers. DNA extracted using this kit is suitable for PCR, qPCR, next generation sequencing (NGS), and other applications.

The GenElute™ UltraMag Cell-Free DNA extraction kit allows preparation of DNA fragments from both fresh and frozen plasma samples. Samples from fresh plasma, however, tends to produce higher yields. Plasma from blood collected with Streck Cell-Free DNA BCT tubes must go through a Proteinase K treatment prior to cell-free DNA isolation to ensure optimal yields. Forgoing Proteinase K treatment may decrease yields by 50%.

The kit provides high DNA recovery with a short and scalable magnetic bead based protocol and without the use of phenol chloroform.

Materials Provided

For 100 mL kit

Item	Catalog No.	Quantity
Lysis/Binding Buffer	CS222724	1 x 115 mL
Wash Buffer	CS222723	2 x 55 mL
Elution Buffer	CS222772	1 x 6 mL
Magnetic Bead Solution	CS222756	2 x 1.33 mL

For 250mL kit.

Item	Catalog No.	Quantity
Lysis/Binding Buffer	CS222770	3 x 95 mL
Wash Buffer	CS222723	5 x 55 mL
Elution Buffer	CS222768	1 x 15 mL
Magnetic Bead Solution	CS222756	5 x 1.33 mL

Materials Required But Not Supplied

Reagents

- 100% ethanol (Sigma Cat. No. E7023 or equivalent)

Samples collected in a Streck Cell-Free DNA BCT tube require:

- Proteinase K (Sigma Cat. No. P2308 or equivalent)
- 20% SDS, (Sigma Cat. No. 05030 or equivalent)

Equipment

- Vortex mixer
- Magnetic stand for molecular applications (Millipore Cat. No. LSKMAGS08, LSKMAGS15 or equivalent)
- 1.5 mL low binding PCR approved vials
- Reaction tubes 2, 15 or 50 mL
- Waterbath or heaterblock

Storage and Stability

Store all kit components at room temperature; performance guaranteed for 1 year from date of receipt when reagents are stored properly.

Protocol

I. Prior to Use

The Lysis/Binding and Wash Buffer are shipped as a concentrate. If precipitate is present in either solution, incubate solution at 37°C for 30 minutes. 100% ethanol must be added to both solutions before the first use. Once ethanol is added, these buffers are stable for one year if stored at 4°C. Be sure to close the bottle tightly for long term storage.

For 100 mL kit

- Add 23 mL of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently
- Add 51 mL of fresh 100% ethanol to each bottle of Wash Buffer and mix by inverting gently
- Prepare fresh 80% ethanol solution prior to each extraction

For 250 mL kit

- Add 19 mL of fresh 100% ethanol to each bottle of Lysis/Binding Buffer and mix by inverting gently
- Add 51 mL of fresh 100% ethanol to each bottle of Wash Buffer and mix by inverting gently
- Prepare fresh 80% ethanol solution prior to each extraction

Before starting the protocol, determine the amount of plasma to be used for extraction and calculate the amount of buffer and beads needed. Any amount from 100 μ L to 10 mL of plasma can be used. Scale buffer and bead volumes accordingly using the table below.

Plasma	Lysis/Binding Buffer	Bead Solution	Tube(s) Size
500 μ L	750 μ L	12.5 μ L	2 mL
1 mL	1.25 mL	25 μ L	15 mL
5 mL	6.25 mL	125 μ L	15 mL or 50 mL*
10 mL	12.5 mL	250 μ L	50 mL

*Using 50 mL tubes for 5 mL or more of plasma is recommended over 15 mL tubes. While using a 15 mL tube will work it may lead to slightly lower yields.

II. Proteinase K Treatment

If samples were collected using a Streck Cell-Free DNA BCT tube(s), Proteinase K treatment is required to ensure optimal yields. If blood was not collected with Streck Cell-Free DNA BCT tube(s) proceed to the Lysis/Binding step.

Plasma	Proteinase K	20% SDS Solution
500 μ L	7.5 μ L	25 μ L
1 mL	15 μ L	50 μ L
5 mL	75 μ L	250 μ L
10 mL	150 μ L	500 μ L

- a. Add the appropriate amount of plasma to an appropriately sized tube(s).
- b. Add 15 μ L of Proteinase K (20 mg/ml) for every 1 mL of plasma used.
- c. Add 50 μ L of 20% SDS solution for every 1 mL of plasma used.
- d. Mix by inverting gently 5 times.
- e. Incubate at 60°C for 20 minutes.

- f. After incubation place tube(s) on ice for 5 minutes to cool tube(s) to room temperature.
- g. Once tube(s) are at room temperature proceed to step 2 of the Lysis/Binding section.

III. Lysis and Binding

- a. Add the appropriate amount of plasma to appropriately sized tube(s).
- b. Add 1.25 mL of Lysis/Binding Buffer for every 1 mL of plasma used.
- c. Add 25 μ L of thoroughly mixed Magnetic Bead Solution for every 1 mL of plasma.

Important: Mix beads well prior to adding by vortexing briefly. There should be no visible sedimentation at the bottom of the solution after mixing. Beads will settle quickly, so be sure to mix the magnetic bead solution after adding to each sample. Failure to do so may result in inconsistent yields.

- d. Vortex or shake tube(s) vigorously for 10 minutes at room temperature.
 - * To obtain high yields, ensure that plasma/buffer solution is mixing vigorously in tube(s). A vortexing mixer with a tube(s)-holder that allows for walk-away mixing will make this easier.
- e. Place tube(s) into a magnet stand for 2 to 5 minutes, or until solution clears.
- f. While keeping the tube(s) on the magnet stand, remove supernatant. Be careful not to remove magnetic particles.
- g. Keep tube(s) on magnet stand for 1 minute, and remove residual supernatant.

IV. First Wash

- a. Add 1000 μ L of Wash Buffer to lysis/binding tube(s).
- b. Resuspend beads by vortexing for 10 seconds or pipetting up and down 6 times.
- c. Transfer magnetic particle suspension into 1.5 mL microcentrifuge tube(s) on magnet stand.
- d. Allow beads to attach to magnet stand for 10-30 seconds, or until solution clears.
- e. Pipette supernatant from 1.5 mL tube(s) and use the supernatant to wash the lysis/binding tube(s).
- f. Transfer the rest of the magnetic particles in lysis/binding tube(s) to the 1.5 mL tube(s).
- g. Keep tube(s) on magnet stand for 10-30 seconds or until solution is clear.
- h. Remove as much buffer as possible using a 1000 μ L pipette.
- i. Tap magnet stand on bench 5 times and remove remaining wash buffer with 200 μ L pipette.
- j. Transfer tube(s) to non-magnetic rack and add 1000 μ L of Wash Buffer.
- k. Resuspend beads by vortexing for 20 seconds or pipetting up and down 6 times.
- l. Centrifuge tube(s) briefly.

*Centrifuge steps are needed when using vortexing to resuspend beads. Only a brief spin is recommended to remove solution from tube(s) lid.

- m. Place tube(s) on magnet stand for 10-30 seconds.

- n. Remove as much buffer as possible using a 1000 μ L pipette.
- o. Tap magnet stand on bench 5 times and remove remaining wash buffer with 200 μ L pipette.

V. Second Wash

- a. Transfer tube(s) to non-magnetic rack and add 1000 μ L of 80% ethanol.
- b. Resuspend beads by vortexing for 20 seconds or pipetting up and down 6 times.
- c. Centrifuge tube(s) briefly.
- d. Place on magnet stand for 10-30 seconds or until solution clears.
- e. Remove as much buffer as possible using a 1000 μ L pipette.
- f. Tap magnet stand on bench 5 times and remove remaining ethanol with 200 μ L pipette.
- g. Transfer tube(s) to non-magnetic rack and add 1000 μ L of 80% ethanol.
- h. Resuspend beads by vortexing for 20 seconds or pipetting up and down 6 times.
- i. Centrifuge tube(s) briefly.
- j. Place on magnet stand for 10-20 seconds.
- k. Remove as much ethanol as possible using a 1000 μ L pipette and leave cap open.
- l. Tap magnet stand with tube(s) on bench 5 times.
- m. Remove remaining ethanol with 200 μ L pipette.
- n. Leave tube(s) open on magnet stand for two minutes and then tap tube(s) on bench 5 times and remove any remaining ethanol with 20 μ L pipette.
- o. Allow magnetic particles to dry for an additional 1-3 minutes. Be careful to not over dry or beads may stick to tube(s).

VI. Elution

- a. Transfer microcentrifuge tube(s) to non-magnetic rack and add desired volume of Elution Buffer and resuspend beads.

Important: A minimum of 12.5 μ L of Elution Buffer per mL of plasma is recommended to elute DNA to ensure optimal yields.

- b. Vortex or shake tube(s) vigorously for 5 minutes.
- c. Centrifuge tube(s) briefly.
- d. Place tube(s) on magnetic rack for 10 to 30 seconds.
- e. Transfer elute into a new 1.5 mL tube(s).

VII. Important Notes

Quantification: Plasma will yield 1-100 ng of Cell-Free DNA per mL of plasma. Therefore, quantification by absorbance measurement (e.g. Nanodrop) may not be sensitive enough to accurately determine yield. Instead, we suggest using QubitTM dsDNA High Sensitivity Assay.

Recommendations for PCR: Due to the highly fragmented nature of the nucleic acids obtained from plasma, care should be taken in the design of primers. Cell-free DNA tends to have a small size (~170bp). Therefore, PCR primers should be designed to produce amplicons of 150 bp or less. Given the low concentration of cfDNA in plasma taken from healthy individuals, 40 amplification cycles may be needed in some cases.

Troubleshooting

Observation	Potential Problems	Experimental Suggestions
Lower than expected yield	Sample degradation during storage.	Repeated freezing and thawing should be avoided. Minimize time between blood draw and plasma preparation if possible.
	An insufficient amount of DNA Magnetic Beads were added.	Vortex the Magnetic Bead Solution immediately before use.
	Magnetic beads are difficult to resuspend after drying step.	Smaller volumes require less time to dry. Airflow and humidity of the immediate environment may change the optimal bead drying times
		Overdried beads will stick to the wall of the plastics and be difficult to re-suspend.
	Magnetic Beads are not completely dried.	Beads that have not been completely dried may carry ethanol into the eluate and impact downstream applications.
	The sample contains low levels of cfDNA.	Increase the volume of the starting sample.
Magnetic Bead carryover	Loose beads are present in the eluate.	Leave the tube on the magnetic stand when removing the eluate containing the cfDNA.
		If beads are carried over into the new tube, place the tube on the magnetic stand again. Wait for the beads to pellet and then transfer the sample into another tube.

Warranty

EMD Millipore Corporation (“EMD Millipore”) warrants its products will meet their applicable published specifications when used in accordance with their applicable instructions for a period of one year from shipment of the products. **EMD MILLIPORE MAKES NO OTHER WARRANTY, EXPRESSED OR IMPLIED. THERE IS NO WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE.** The warranty provided herein and the data, specifications and descriptions of EMD Millipore products appearing in EMD Millipore’s published catalogues and product literature may not be altered except by express written agreement signed by an officer of EMD Millipore. Representations, oral or written, which are inconsistent with this warranty or such publications are not authorized and if given, should not be relied upon.

In the event of a breach of the foregoing warranty, EMD Millipore Corporation’s sole obligation shall be to repair or replace, at its option, the applicable product or part thereof, provided the customer notifies EMD Millipore Corporation promptly of any such breach. If after exercising reasonable efforts, EMD Millipore Corporation is unable to repair or replace the product or part, then EMD Millipore shall refund to the Company all monies paid for such applicable Product. **EMD MILLIPORE CORPORATION SHALL NOT BE LIABLE FOR CONSEQUENTIAL, INCIDENTAL, SPECIAL OR ANY OTHER DAMAGES RESULTING FROM ECONOMIC LOSS OR PROPERTY DAMAGE SUSTAINED BY ANY COMPANY CUSTOMER FROM THE USE OF ITS PRODUCTS.**

Unless otherwise stated in our catalog or other company documentation accompanying the product(s), our products are intended for research use only and are not to be used for any other purpose, which includes but is not limited to, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses or any type of consumption or application to humans or animals.

(c) 2009 - 2014: Merck KGaA, Darmstadt, Germany. The M mark is a registered trademark of Merck KGaA, Darmstadt, Germany. All rights reserved. No part of these works may be reproduced in any form without permission in writing.

Catalog No. CFMAG

01 2018
Revision A: GKSMAN