


# 50 – 1000 µl Pipette tips 732012 | 732032




Dimensions in mm

<b>Volume</b>	50 – 1000 µl
<b>Material</b>	PP; autoclavable at 121 °C (2 bar) acc. to DIN EN 285
<b>Color</b>	blue
<b>Packaging</b>	bulk
<b>Quality</b>	Produced under cleanroom conditions. CE-marked acc. to EC-Directive 98/79 EC (IVD).

# 50 – 1000 µl Pipette tips

## 732112 | 732212 | 732252



Dimensions in mm


<b>Volume</b>	50 – 1000 µl
<b>Material</b>	PP; autoclavable at 121 °C (2 bar) acc. to DIN EN 285
<b>Color</b>	transparent
<b>Packaging</b>	racked
<b>Quality</b>	Cleanroom class 8 according to ISO 14 644-1, Free of DNA*, RNase and DNase, Free of pyrogens, according to LAL test **, CE-marked acc. to EC-Directive 98/79 EC (IVD).



\* human and bacterial DNA, \*\* according to Limulus Amebocyte Lysate (LAL) test, detection limit 0.01 EU/ml

# 50 – 1000 µl Pipette tips

## 732132 | 732232 | 732272



Dimensions in mm

<b>Volume</b>	50 – 1000 µl
<b>Material</b>	PP; autoclavable at 121 °C (2 bar) acc. to DIN EN 285
<b>Color</b>	transparent
<b>Packaging</b>	racked
<b>Quality</b>	Cleanroom class 8 according to ISO 14 644-1, Free of DNA*, RNase and DNase, Free of pyrogens, according to LAL test **, Sterile according to ISO 11 137, Free of ATP ,CE-marked acc. to EC-Directive 98/79 EC (IVD).



\* human and bacterial DNA, \*\* according to Limulus Amebocyte Lysate (LAL) test, detection limit 0.01 EU/ml