



Servicebio

产品检验报告

Product test report

产品名称 Product name	移液管 Serological Pipette		
产品批号 Lot. No	04922136	产品材质 Material	PS (聚苯乙烯) Polystyrene
产品型号 Product model	\	产品规格 Specs	1ml
产品货号 Cat. No	SP-1-SF	储存条件 storage	常温 Room temperature
产品状态 Status	无菌 无热原 Dnase/ Non-pyrogenic	有效期 Validity	2 年 2 years
检验依据:《产品质量检验标准》、GB/T 2828.1 Test basis: Pharmacopoeia of the People's Republic of China 2020 edition, GB/T 2828.1-2012			
检验项目 Test items			
检验指标 Test index	检验标准/依据 Test standard	检验结果 Test result	判定 Decision
产品外观 Appearance	迎光距离 30cm 观察: 孔是白色半透明、无油污、无杂质、无毛刺、无明显划痕; 四周无开裂、无毛刺、无变形。 Observation at a 30cm front light distance: The hole is white transparent, no oil, no impurities, no burrs, no obvious scratches; No cracks, no burrs, no deformation around the plates.	符合要求 All the products are qualified	合格 Qualified
尺寸 Size	关键尺寸(长、宽、高)符合图纸要求。 The key dimensions meet the requirements of the drawings.	符合要求 All the products are	合格 Qualified
功能测试 Functional testing	吸排自如, 不允许有渗漏; 刻度精准, 允许误差 $\pm 2\%$ Suction and discharge freely, no leakage is allowed; Accurate calibration, allow error $\pm 2\%$.	符合要求 All the products are qualified	合格 Qualified
无菌检验 Sterile test	细菌培养阴性 Bacteria culture negative.	阴性 Negative	合格 Qualified
热原 Pyrogenic	《中华人民共和国药典》2020 版小于 0.5EU/mL: 检测样品均阴性 Pharmacopoeia of the People's Republic of China 2020 version less than 0.5EU/mL: All the tested samples were neg	阴性 Negative	合格 Qualified
检验结果 Test result: 根据《产品质量检验标准》中检验项目进行检验, 所有检验项目均符合要求, 因此此批产品检验合格, 可入库发货。 According to the inspection items in the product inspection reference standard, all inspection items meet the requirements. Therefore, this batch of products can be stored and delivered.			
 质检专用章 Special stamp for quality inspection 签发日期 Issuing date: 2022.02.28			

检验: 张赛、桂良凤 审核: 李子庆 批准: 刘小影 表单编号: SWE-QA-SMP-006-03
 Inspected by: Zhang Sai、Gui Liangfeng Checked by: Li Ziqing Approved by: Liu xiaoying No: SWE-QA-SMP-006-03